



Field Summary Report:	September 14 th , 2023
Client Name:	CAMCO
Inspector(s):	Kevin Loftus
Inspection Timeframe:	9:30am → 10:30am EST

Inspection Summary:

- A total of **2x** AEDs were fully inspected, interrogated, and updated on September 14th, 2023.
- AEDs reported NO Error Codes.
- AEDs PASSED all Daily/Weekly/Monthly Self-Tests.
- AEDs were returned to their respective Storage Locations in **Response Ready** condition and certified.
- All AEDs are compliant with all Pennsylvania State and Federal AED laws and regulations until October 2024.
- All AEDs are certified with new inspection tags.
- **Parts Replaced/Updated Totals:**
 - Adult Pads Replaced: N/A
 - Pediatric Pads Replaced: N/A
 - Batteries Replaced: N/A

****Serial Numbers Inspected, Updated, and Certified with NO Concerns****

Zoll AED Plus

1. Serial Number: **X23B587417** located @ 608 Regency Hill Drive, Collegeville, PA 19426
2. Serial Number: **X23C588909** located @ 199 Sloan Road, Phoenixville, PA 19460

Recommendations:

- I **HIGHLY** recommend scheduling AED In-Service Training Sessions included in the cost of the agreement. Training is highly beneficial and designed to build confidence in anyone who may be called upon to deploy an AED during an emergency. Trainings are delivered virtually via the Zoom Platform. **We can schedule as many virtual Training Sessions as are necessary to ensure all who wish to attend are able.**

To schedule In-Service Training Sessions, please contact Nate Marin at NM@AEDserviceAmerica.com OR 860-331-6975



Status of Deployment:

1. Serial Number: **X23B587417** located @ 608 Regency Hill Drive, Collegeville, PA 19426
 - a. Primary Pads Expire: 10/28/2027
 - b. Battery Expires: 5/11/2028

2. Serial Number: **X23C588909** located @ 199 Sloan Road, Phoenixville, PA 19460
 - a. Primary Pads Expire: 11/6/2027
 - b. Battery Expires: 5/11/2028